

## AGSS Receiving System

SPECIFICATION SHEET

The AGSS is designed to conform to ISO 8835-3 and is used for the extraction of waste and excess gases, from the patient breathing system into the central anaesthetic gas scavenging disposal system.

This is a device that safely collects and removes exhaled nitrous oxide and other anaesthetic gases during the administration of anaesthesia or sedation. The device is recommended by: AAGBI to avoid occupational exposure to the gases.

- Conforms to ISO 8835-3
- MRI Compatible
- Wall, rail or machine mounted
- Weighs 1 Kg, size 80mm x 390mm
- Supplied with Disposal and Transfer hoses
- Both Disposal Hose and Patient Inlet port swivel about 360° to ensure hose positioning and prevent occlusion.



Supplied by

**MEC**  
MEDICAL LIMITED

MEC House, Ivinghoe Business Centre, Blackburn Road,  
Houghton Regis, Bedfordshire. LU5 5BQ. United Kingdom

T: +44 (0) 1582 661885 F: +44 (0) 1582 602527  
sales@mecmedical.com www.mecmedical.com